

Notes:

- Still working on direct terminal
- Two spi issue
 - We think it is a bluetooth issue
 - Direct SPI communication works just fine
 - Need to figure out what in the software is going out
- Water tests
 - Sponge test went well for the longer wavelengths
 - Varies based on the position of the individual being tested
 - How the blood is flowing (ie blood viscosity)
- Demo day
 - Sponge testing
 - Maybe cuvette
 - Working on a human would be prime